

## Laser Power Reading

**Microscope:** NOBIC custom DNA damage microscope

**Date:** 13/5/2026

**Objective:** 20x/0.45 Air

**Parameters:** 405 nm

Setting on laser display	Power at sample plane / $\mu$ W
78	8000
70	6600
60	4200
50	2200
40	180
30	5
20	1.8
10	0.6